

Tender No:

UHKDU/PRO/MED/111/2020



University Hospital Kotelawala Defence University  
Werahara, Boraesgamuwa  
Sri Lanka  
T:Phone: 0112044555  
Fax:  
Web:

29/09/2020

## **INVITATION TO BID AND GENERAL CONDITIONS OF TENDER**

1. The Vice Chancellor of the General Sir John Kotelawala Defence University, as the Chairman, Department Procurement Committee invites Bid/s from prospective Bidders for supply of **items listed in the schedule in Annex "A" The relevant specifications of the items are indicated in Annex "B"**

2. **CLOSING DATE & TIME.** The tender will close at 1400 hrs on **21 / 10 / 2020** Any Bid submitted after the closing time of the tender will be rejected & unopened such bids will be returned to the bidder.

3. **VALIDITY OF BID.** The bid submitted under this tender **must be valid for a period of 120 days from the date of closing of tender.**

### **BID BOND / GUARANTEE.**

(a) If the quoted bid value exceeds Rs: 2,000,000.00, such Bids should be accompanied with 1% of Bid Value an "on demand" and "unconditional" Bid Bond/Guarantee for a sum of Rs: ..... in the format given in Annex "C" through a recognized local Bank or Insurance Company registered in Sri Lanka which is authorized by the Insurance Board of Sri Lanka to issue such Bid Guarantees. All Bid Bond/Guarantees should be valid for at least 30 days more than the validity period of bids, ie, for 150 days from the date of Bid opening. **Cheques will not be accepted as Bid Guarantee.**

(b) Submission of insufficient Bid Bond/Guarantee value or period will be considered as a "**major deviation**" and such offer will not be considered for further procurement action and will be rejected.

5. **VALUE ADDED TAX.** The Bidders who bid for locally delivered items must have the VAT registration. The VAT portion must be shown separately in the price schedule in Annex "A" and VAT registration number must be indicated. If the quoted item is exempted from VAT or Bidding Company is not liable for VAT, reference number and date of relevant Act number/Gazette notification/a certificate (as applicable) **issued for the current financial year** from the Commissioner General of Inland Revenue to that effect should be submitted along with the Bid.

6. **BID SUBMISSION.** The **bidder must duly sign at the last page (before Annexes) of this document** indicating the name of the signatory and the name of the company & place the company common seal to confirm the acceptance of tender conditions. **The Bid/s that do not include authorized signature will be rejected.** The Bid/s duly signed by the bidder enclosed and sealed in an appropriate cover addressed to the following address should be sent by registered post or could be deposited in the **appropriate tender box placed at University Hospital Kotelawala Defence University, Werahara** (at the Main Entrance of University Hospital Kotelawala Defence University) on or before the time & date specified for the closing of tender. **The tender reference number, date & time of closing tender should be indicated & underlined at the top left corner of the envelop.**

**Tender Ref No** .....  
**Closing Date & Time** .....

The Chairman,  
Department Procurement Committee,  
University Hospital Kotelawala Defence University,  
Werahara,  
Sri Lanka.

7. The Bid/s must be submitted in the attached schedule of prices in Annex "A" as applicable. However, bidders could use similar formats prepared with their own letter heads with all the details mentioned therein and submit in **three copies along with duly signed copy of a General Conditions of tender.** The Bid/s must contain **Technical Literature, Pamphlets, Drawings and Quality Standard Certificate etc** necessary to determine characteristics of items offered and in case of Machinery/Vehicles & Equipment, servicing and workshop data/after sales service, back up facilities or any other facilities provided by the supplier.

### **SUBMISSION OF SAMPLES/PAYMENT OF TESTING CHARGES.**

When it is required to submit samples, **every offer** must be accompanied with pre - marked samples. The marking of samples **indicating the Bidder & Offer number** must be done and the samples must be handed over to the officer at same place where tender box is placed **on or before the closing date & time of the Bid.** Samples submitted after closing time of the Bid will be rejected. The documents such as Air Way Bills etc, will not be accepted in place of samples. When the testing charges are required to be paid, bidders shall pay testing charges separately **for all offers** indicated in their bid/s.

(1) **Samples.** Please submit ..... samples.

(2) **Testing Charges.** A sum of Rs. .... **per offer** must be paid to the Bursar of KDU, prior to the submission of bid and a copy of the receipt must be annexed to the bid. It is the responsibility of bidder to inform the Cashier of the Account Office to note the tender number on the receipt issued for such payments.

8. **BID OPENING.** All duly received bids **will be opened immediately after the scheduled closing time of Bids at the same venue.** Bidders or their accredited agents could be present at the time of opening of bids.

### **PRICES.**

For locally delivered items (including locally manufactured items & foreign items imported by the bidders) price must be quoted in Sri Lankan Rupees, inclusive of all charges for delivery of items to University Hospital General Sir John Kotelawala Defence University Werahara. **Unit price, VAT and Total price should be clearly indicated in schedule in Annex "A".** Other than VAT, all other type of taxes (eg : NBT, BTT, etc.) should not be indicated separately and should be included in unit price.

10. **RESTRICTED TENDERS.** Invitation to Bids are circulated among the registered suppliers with Ministry of Defence (MOD), only bids submitted by registered suppliers will be allowed for consideration. However, Chairman, Department Procurement Committee reserves the right to invite the bids from multiple combinations of Procurement Methods as stipulated in Chapter III of the Government Procurement Guideline, 2006 to ensure highest competitiveness.

11. **PERFORMANCE BOND/GUARANTEE.** A successful bidder shall furnish a Performance Bond/Guarantee in the form of "On Demand" & "Unconditional" Bank/Insurance Guarantee for a sum equivalent to 10% of the contract value for every contract that exceeds Rs. 500,000.00 or equivalent amount in foreign currency through a recognized Commercial Bank registered in Sri Lanka through an Insurance Company authorized by the Insurance Board of loss resulting from the supplier's failure to complete his performance obligations under the contract. If the contracted supplier fails to deliver the items Sri Lanka to issue such Performance Guarantee for this purpose, within two weeks from the date of notification of award. The proceeds of the Performance Bond/Guarantee shall be payable to the Vice Chancellor of the General Sir John Kotelawala Defence University as compensation for any o

12. time or fails to complete the works as per the agreed contract, **THE TOTAL VALUE OF THE PERFORMANCE BOND/GUARANTEE** will be forfeited. If only partial delivery is made during the agreed contract period, the corresponding value percentage of undelivered quantity from the Performance Bond/Guarantee will be forfeited.

13. **SIGNING OF CONTRACT.** The notification of award will be transmitted to the successful bidder by post, by fax or e-mail. This notification constitutes the formation of the contract. The successful bidder should submit his written acceptance for the award and performance bond / guarantee (For awards over Rs. 500,000.00 without VAT) within 14 days of receipt of such notification. Upon acceptance of the award and furnishing of the Performance Bond/Guarantee, the successful bidder will have to enter into a formal contract with the Vice Chancellor of the General Sir John Kotelawala Defence University by signing the Contract.

14. **DELIVERY.** Preference will be given for early delivery. In case of bulk supplies for locally manufactured items, the delivery of **total quantity must be completed within 120 days of signing of contract**, unless mutually agreed for extended delivery period with University Hospital Kotelawala Defence University. **The bidder/s must indicate the proposed delivery schedule in Annex "E"**. In the event of placing a purchase order with the successful bidder, the total quantity so ordered must be supplied as one consignment unless part deliveries are agreed upon in the contract. The deliveries not made as per agreed delivery schedule will be considered as bad performances by the suppliers and **no extended delivery period will be authorized**. Under extreme unavoidable conditions too, the Chairman, Department Procurement Committee (Vice Chancellor of the General Sir John Kotelawala Defence University) reserves the right to grant or refuse delivery period extensions only within the current financial year with or without liquidated claim for delayed deliveries and that decision will be final.

15. **LIQUIDATED DAMAGES.** In case of delivery period extensions requested by the successful bidder, a sum equivalent to 2% of the total value of the delayed supply per delay of one week or part thereof may be deducted from the payment due to the supplier from the University Hospital General Sir John Kotelawala Defence University as liquidated damages up to the maximum limit of 10% of the total value of delayed supplies.

16. **PAYMENT TERMS FOR LOCALLY DELIVERED ITEMS.** Payment will be made after acceptance of items which should be subjected to a pre-acceptance inspection/testing by University Hospital General Sir John Kotelawala Defence University authorities. The delivery made to University Hospital should not be considered as quantities taken over by University Hospital General Sir John Kotelawala Defence University until items are properly accepted after pre-acceptance inspection. Any item that does not conform to the specifications or already approved sample will be rejected & it is the responsibility of the supplier to remove them from University Hospital stores/premises within 07 working days of such intimation (either verbal or written) at his own cost and replace them with items conforming to specification within one month of such rejection. The bidder shall allow approximately 60-90 days period of credit from the date of acceptance of items for Account Office, University Hospital General Sir John Kotelawala Defence University to obtain liquid cash from General Treasury & release the payment.

16. **RIGHTS OF THE PROCUREMENT COMMITTEE.** The Department Procurement Committee reserve the right to accept or reject whole or part of this tender and their decision will be final. The successful bidders will be notified. Information with regard to rejected or unsuccessful bids will not be communicated.

Thanking You,  
Yours faithfully,



For Chairmen DPC  
General Sir John Kotelawala Defence University  
University Hospital Kotelawala Defence University

I/We agree to abide by the conditions of tender and undertake to supply the items as per delivery schedule mentioned in the contract, in the event of an order been placed with me/my firm/company as a result of this tender.

..... Signature  
..... Name of Signatory

.....Name of the Company/Bidder

Date :- .....

Company seal

**SCHEDULE OF PRICES FOR LOCALLY DELIVERED ITEMS**

S/ N	ITEMS	UNIT	QTY	PRICE EACH SLRS	TOTAL PRICE SLRS
	<b><u>PURCHASE OF MEDICAL ACCESSORIES FOR 02 NO'S AMBULANCE AT UHKDU</u></b>				
01	Portable Multi Parameter Monitor AC/DC	No's	02		
02	Portable Transport Ventilator (Adult/Pediatric) AC/DC	No's	02		
03	Portable Nebulizer AC/DC	No's	02		
04	Portable Suction until AC/DC with Provisions for Pedal Operation	No's	02		
05	Reusable Ambu Bags (Adult/Pediatric)	No's	02		
06	Oxygen Flow Meter with Regulator	No's	02		
	DISCOUNT				
	TOTAL(AFTER DISCOUNT)				
	VAT %				
	<b>GRAND TOTAL</b>				

2. a. Country of Manufacture :-.....

b. Name of the manufacturer and address :- .....

3. VAT Registration No.: .....

4. Name of the Company : .....

5. Testing Charges Paid. : Rs..... Receipt No and Date: .....

6. Signature of Bidder : .....

7. Date. : .....

8. MOD registered No. : ..... Company Seal :- .....

## Technical Specifications

## IN-BUILT MEDICAL ACCESSORIES FOR 02 NO'S AMBULANCE

Description: 02 No's Potable Multi Parameter Monitor AC/DC

1	2	3	4	5	6	
	KDU Requirements	Priority <sup>2</sup>	Conformity		Proof <sup>4</sup>	Bidder's Offer Remarks <sup>3</sup>
			Yes	No		
1	<b>Potable Multi Parameter Monitor AC/DC</b>					
	i. Monitors shall be modular and supplied with all necessary fixtures and wall mounts	C				
	ii. Shall have configuration setting to adjust the monitor for adult & paediatric patient applications	C				
	iii. The system shall have alarm indicators, which are graded and prioritized according to their severity and should be both available in visual and audible from	C				
	iv. All transducer input connectors shall be color/key coded	C				
	v. Power supply requirements: The operation voltage for all hardware shall be derived from AC main supply of 220-230V, 50Hz and DC for Ambulance	C				
	vi. Monitors shall have two(2) hours battery back up	C				
	vii. Display	C				
	a) Shall be high resolution color type with different color coded traces and shall measure minimum of 12 inches diagonally					



	<p>b) The display shall have a bright screen, easy viewable even at a distance from any angle under all lighting conditions</p> <p>c) Should have capability to display at least 4 real time waveforms along with related numerical parameters on a digital screen</p> <p>d) The monitors should have the capability to operate through selectable user interface with touch screen</p> <p>e) Shall have facility to freeze wave forms</p> <p>f) Shall have different display settings</p>				
	<p>viii. Parameters: Monitors shall be able to measure following parameters simultaneously</p> <p>a) 3/5 lead ECG</p> <p>b) None invasive blood pressure (NIBP)- anaeroid type</p> <p>c) Pulse oxymetry –SpO2 (motion tolerant wave from)</p>	C			
	<p>ix. Accessories: a) Adult/paediatric size BP cuffs</p> <p>b) Pulse oximeter probes(adult/paediatric sizes)</p>	C			
2	<b>Power</b>	C			
	<p>i. AC voltage (240V+/-10%).</p> <p>ii. DC (12V+/-10% ambulance power) or supplier have to provide suitable DC to AC inverter free of charge.</p>	C			
	<p>iii. Battery Backup (maintenance-free) Minimum 60 minutes on full charge.</p> <p>iv. Should be provided with AC adaptor (3-pin square hospital grade plug top) and cigarette lighter socket cable for use in ambulance.</p>	C			
3	All items supplied should be in-built in ambulance	C			

4	<b>Warranty, Services Agreements and Documentation</b>	C					
	i. All items described in this schedule should be covered by a warranty for a period of not less than 24 calendar months from the date of successful commissioning of the equipment including for any defect in manufacturing process, workmanship and all accessories	C					
	ii. A Labour-only service contract for a 03 year period and comprehensive maintenance contract (CMC) amount for thr whole equipment including installation material for another 5 years should be quoted as per schedule given below after 02 year warranty period	C					
	a) 1 <sup>st</sup> & 2 <sup>nd</sup> year : warranty period						
	b) 3 <sup>rd</sup> to 5 <sup>th</sup> year : Labour only contract	C					
	c) 6 <sup>th</sup> to 10 <sup>th</sup> year : Comprehensive maintenance contract	C					
	iii. The technology manuals having full details of installation, operations, service and maintenance in the English language must be made available in duplicate at the time of installation and commissioning	C					
	iv. Spares/replaceable items/consumables required for the system must be guarantees for supply over a period of 10 years from the date of installation. The price list for such items over a period of 5 years from the date of supply should be made available	C					
	v. A sample unit must be available for demonstration and performance evaluation at the request of TEC	C					
	vi. Prospective suppliers must provide documentary evidence of their previous experience with similar units installed in Sri Lanka as well as evidence of the reliability of after-sales services provided.	C					
	vii. Training should be provided to end users in respect of operation, handling and maintenance of equipment supplied	C					

(2)

Make & Model: .....  
Country of Manufacture: .....

Relevant catalogues & technical information attached.  
Information given under "Bidder's Offer" is true and correct.

Bidder's Name, signature & Date  
Company seal

- 
- <sup>1</sup> Bidder shall fill in Col. 4, 5 & 6
  - <sup>2</sup> Purchaser will mark C for critical requirements
  - <sup>3</sup> If the bidder has marked No for Conformity, bidder shall provide information in this column of his offer against the clause concerned in col. 2 (Purchaser's requirements). If the bidder has marked Yes, Bidder has the option of providing additional information to establish that it conforms to the specification given
  - <sup>4</sup> Bidder shall include proof to support the conformity of the specification e.g. document and the page number in the original user/service manual mentioning the specification.





	e) PEEP/CPAP – 0 to 30 cm H <sub>2</sub> O								C
	f) PS (Pressure support) - 0 to 50 cm H <sub>2</sub> O								C
	g) Pressure trigger (Sensitivity) - -9.9 to 0 cm H <sub>2</sub> O								C
	h) Flow trigger – 0.0 to 10.0 L/min								C
	i) Flow – 6 to 100 L/min								C
	j) SIMV rate – 6-30 bpm								C
	k) I:E Ratio – 1:99 to 3:1								C
	l) FIO <sub>2</sub> – 21 to 100%								C
	m) Flow wave pattern – Square or descending								C
	vi. Monitored Parameters:								C
	Paw (Peak, Mean,Base) – insp/exp tidal volume								C
	Insp/exp minute volumes – Peak insp. Flow								C
	O <sub>2</sub> Concentration – I:E Ratio								C
	vii. Alarms:								C
	High/Low Airway pressure – Apnea								C
	High/Low Mean pressure – High/Low Oxygen								C
	High Resp. Rate – Low Battery								C
	High Tidal Volume – Airway flow sensor disconnect								C
	viii. Patient presets: Customized or default								C
	ix. Customized back up ventilation: Link with apnea and/or low MV and customize settings								C
	x. Should be supplied with following accessories:								C
	a) Oxygen supply hose and regulator								C

	a) Adult and paediatric reusable patient circuits (Silicone)	C				
	b) Test lungs (adult/paediatric)	C				
	xi. Shall have following alarms:	C				
	a) Peak pressure	C				
	b) High/low breath rate	C				
	c) High/low Oxygen %	C				
	d) Low battery	C				
	xii. Ventilator should be equipped with a dual heated servo controlled humidifier system with adult paediatric chambers and accessories	C				
2.	<b>Power</b>					
	i. AC voltage (240V+/-10%).	C				
	ii. DC (12V+/-10% ambulance power) or supplier have to provide suitable DC to AC inverter free of charge.	C				
	iii. Battery Backup (maintenance-free) Minimum 60 minutes on full charge.	C				
	iv. Should be provided with AC adaptor (3-pin square hospital grade plug top) and cigarette lighter socket cable for use in ambulance.	C				
4.	All items supplied should be in-built in ambulance	C				
5.	<b>Warranty, Services Agreements and Documentation</b>					
	i. All items described in this schedule should be covered by a warranty for a period of not less than 24 calendar months from the date of successful commissioning of the equipment including for any defect in manufacturing process, workmanship and all accessories	C				
	ii. A Labour-only service contract for a 03 year period and comprehensive maintenance contract (CMC) amount for the whole	C				

	equipment including installation material for another 5 years should be quoted as per schedule given below after 02 year warranty period					
	a) 1 <sup>st</sup> & 2 <sup>nd</sup> year : warranty period					
	b) 3 <sup>rd</sup> to 5 <sup>th</sup> year : Labour only contract	C				
	c) 6 <sup>th</sup> to 10 <sup>th</sup> year : Comprehensive maintenance contract	C				
	iii. The technology manuals having full details of installation, operations, service and maintenance in the English language must be made available in duplicate at the time of installation and commissioning	C				
	iv. Spares/replaceable items/consumables required for the system must be guaranteed for supply over a period of 10 years from the date of installation. The price list for such items over a period of 5 years from the date of supply should be made available	C				
	v. A sample unit must be available for demonstration and performance evaluation at the request of TEC	C				
	vi. Prospective suppliers must provide documentary evidence of their previous experience with similar units installed in Sri Lanka as well as evidence of the reliability of after-sales services provided.	C				
	vii. Training should be provided to end users in respect of operation, handling and maintenance of equipment supplied	C				

Make & Model: .....  
Country of Manufacture: .....

Relevant catalogues & technical information attached.  
Information given under "Bidder's Offer" is true and correct.

Bidder's Name, signature & Date

Company seal

- <sup>1</sup> Bidder shall fill in Col. 4, 5& 6
- <sup>2</sup> Purchaser will mark C for critical requirements
- <sup>3</sup> If the bidder has marked No for Conformity, bidder shall provide information in this column of his offer against the clause concerned in col. 2 (Purchaser's requirements). If the bidder has marked Yes, Bidder has the option of providing additional information to establish that it conforms to the specification given
- <sup>4</sup> Bidder shall include proof to support the conformity of the specification e.g. document and the page number in the original user/service manual mentioning the specification.

Technical Specifications

IN-BUILT MEDICAL ACCESSORIES FOR 02 NO'S AMBULANCE

Description: 02 No's Portable Nebulizer AC/DC

1	2	3	4	5	6	
	KDU Requirements	Priority <sup>2</sup>	Conformity		Proof <sup>4</sup>	Bidder's Offer Remarks <sup>3</sup>
			Yes	No		
4	<b>Portable Nebulizer AC/DC</b>					
	i. The equipment should be hospital grade product, light weight, portable	C				
	ii. Nebulizer shall operate on AC/DC on vehicle battery power (12V)	C				
	ii. Minimum air flow should be not less than 12 liters per minute	C				
	i. The output pressure should be not less than 0.5 bars	C				
	ii. The noise level should be less than 60 dB	C				
	iii. The equipment shall incorporate with ON/OFF switch, line pressure gauge and a carrying handle	C				
2.	<b>Power</b>					
	i. AC voltage (240V+/-10%).	C				
	ii. DC (12V+/-10% ambulance power) or supplier have to provide suitable DC to AC inverter free of charge.	C				
	iii. Battery Backup (maintenance-free) Minimum 60 minutes on full charge.	C				

	iv. Should be provided with AC adaptor (3-pin square hospital grade plug top) and cigarette lighter socket cable for use in ambulance.	C				
3.	All items supplied should be in-built in ambulance	C				
4.	<b>Warranty, Services Agreements and Documentation</b>					
	i. All items described in this schedule should be covered by a warranty for a period of not less than 24 calendar months from the date of successful commissioning of the equipment including for any defect in manufacturing process, workmanship and all accessories	C				
	ii. A Labour-only service contract for a 03 year period and comprehensive maintenance contract (CMC) amount for thr whole equipment including installation material for another 5 years should be quoted as per schedule given below after 02 year warranty period a) 1 <sup>st</sup> & 2 <sup>nd</sup> year : warranty period b) 3 <sup>rd</sup> to 5 <sup>th</sup> year : Labour only contract c) 6 <sup>th</sup> to 10 <sup>th</sup> year : Comprehensive maintenance contract	C				
	iii. The technology manuals having full details of installation, operations, service and maintenance in the English language must be made available in duplicate at the time of installation and commissioning	C				
	iv. Spares/replaceable items/consumables required for the system must be guarantees for supply over a period of 10 years from the date of installation. The price list for such items over a period of 5 years from the date of supply should be made available	C				
	v. A sample unit must be available for demonstration and performance evaluation at the request of TEC	C				
	vi. Prospective suppliers must provide documentary evidence of their previous experience with similar units installed in Sri Lanka as well as evidence of the reliability of after-sales services provided.	C				

vii. Training should be provided to end users in respect of operation, handling and maintenance of equipment supplied	C				
---	---	--	--	--	--

Make & Model: .....  
Country of Manufacture: .....

Relevant catalogues & technical information attached.  
Information given under "Bidder's Offer" is true and correct.

Bidder's Name, signature & Date

Company seal

- <sup>1</sup> Bidder shall fill in Col. 4, 5& 6
- <sup>2</sup> Purchaser will mark C for critical requirements
- <sup>3</sup> If the bidder has marked No for Conformity, bidder shall provide information in this column of his offer against the clause concerned in col. 2 (Purchaser's requirements). If the bidder has marked Yes, Bidder has the option of providing additional information to establish that it conforms to the specification given
- <sup>4</sup> Bidder shall include proof to support the conformity of the specification e.g. document and the page number in the original user/service manual mentioning the specification.

Technical Specifications

IN-BUILT MEDICAL ACCESSORIES FOR 02 NO'S AMBULANCE

Description: 02 No's Portable suction unit AC/DC with provisions for pedal operation

1	2	3	4	5	6
	<b>KDU Requirements</b>				<b>Bidder's Offer</b>
		<b>Priority<sup>2</sup></b>	<b>Conformity</b>	<b>Proof<sup>4</sup></b>	<b>Remarks<sup>3</sup></b>
			Yes	No	
1	<b>Portable suction unit AC/DC with provisions for pedal operation</b>				
	i. The unit should be floor standing on anti-static casters with breaks and suitable for continuous operation	C			
	ii. The unit shall consist of 2 jars (autoclavable) made of polycarbonate with self-sealing polypropylene tubing with a capacity of 2 liters	C			
	iii. The unit should have three way protection at trap bottle and hydrophobic shut off valve filter just before the motor	C			
	iv. Pump type – oil free piston pump (electrical) and piston pump (pedal)	C			
	v. Jar autoclavable with mechanical/electronic over flow motor shut off system and autoclavable bacterial filter tubing(silicon)	C			
	vi. Power-electrical operation on 220V A/C, 6V, 4.5 Amp rechargeable battery in-built pedal	C			
	vii. Battery backup – 120 min on full charge	C			
	viii. Should be provided with AC adaptor and cigarette lighter socket cable for use in ambulance	C			
	ix. The pump should be oil less type with two pistons	C			



80

2	<b>Power</b>						
	i.	AC voltage (240V+/-10%).	C				
	ii.	DC (12V+/-10% ambulance power) or supplier have to provide suitable DC to AC inverter free of charge.	C				
	iii.	Battery Backup (maintenance-free) Minimum 60 minutes on full charge.	C				
	iv.	Should be provided with AC adaptor (3-pin square hospital grade plug top) and cigarette lighter socket cable for use in ambulance.	C				
3		All items supplied should be in-built in ambulance	C				
4		<b>Warranty, Services Agreements and Documentation</b>					
	i.	All items described in this schedule should be covered by a warranty for a period of not less than 24 calendar months from the date of successful commissioning of the equipment including for any defect in manufacturing process, workmanship and all accessories	C				
	ii.	A Labour-only service contract for a 03 year period and comprehensive maintenance contract (CMC) amount for thr whole equipment including installation material for another 5 years should be quoted as per schedule given below after 02 year warranty period a) 1 <sup>st</sup> & 2 <sup>nd</sup> year : warranty period b) 3 <sup>rd</sup> to 5 <sup>th</sup> year : Labour only contract c) 6 <sup>th</sup> to 10 <sup>th</sup> year : Comprehensive maintenance contract	C				
	iii.	The technology manuals having full details of installation, operations, service and maintenance in the English language must be made available in duplicate at the time of installation and commissioning	C				
	iv.	Spares/replaceable items/consumables required for the system must be guarantees for supply over a period of 10 years from the date of	C				

	installation. The price list for such items over a period of 5 years from the date of supply should be made available					
	v. A sample unit must be available for demonstration and performance evaluation at the request of TEC	C				
	vi. Prospective suppliers must provide documentary evidence of their previous experience with similar units installed in Sri Lanka as well as evidence of the reliability of after-sales services provided.	C				
	vii. Training should be provided to end users in respect of operation, handling and maintenance of equipment supplied	C				

Make & Model: .....

Country of Manufacture: .....

Relevant catalogues & technical information attached.  
 Information given under "Bidder's Offer" is true and correct.

Bidder's Name, signature & Date

Company seal

- <sup>1</sup> Bidder shall fill in Col. 4, 5& 6
- <sup>2</sup> Purchaser will mark C for critical requirements
- <sup>3</sup> If the bidder has marked No for Conformity, bidder shall provide information in this column of his offer against the clause concerned in col. 2 (Purchaser's requirements). If the bidder has marked Yes, Bidder has the option of providing additional information to establish that it conforms to the specification given
- <sup>4</sup> Bidder shall include proof to support the conformity of the specification e.g. document and the page number in the original user/service manual mentioning the specification.

Technical Specifications

IN-BUILT MEDICAL ACCESSORIES FOR 02 NO'S AMBULANCE

Description: 02 No's Reusable Ambu Bags (adult/paediatric)

1	2	3	4	5	6
	<b>KDU Requirements</b>	<b>Priority<sup>2</sup></b>	<b>Conformity</b>	<b>Proof<sup>4</sup></b>	<b>Bidder's Offer</b>
			Yes		<b>Remarks<sup>3</sup></b>
			No		
1	<b>Reusable Ambu Bags (adult/paediatric)</b>				
	i. 100% Latex Free	C			
	ii. Autoclavable	C			
	iii. Silicone self inflating bag (both adult and paediatric sizes)	C			
	iv. All parts can be removed for cleaning and disinfecting	C			
	v. Accessories	C			
	a) 2M Oxygen supply line				
	b) Oxygen reservoir	C			
	c) Silicone reusable face marks – both adult/paediatric sizes (3 sizes from each)	C			
2	<b>Power</b>				
	i. AC voltage (240V+/-10%).	C			
	ii. DC (12V+/-10% ambulance power) or supplier have to provide suitable DC to AC inverter free of charge.	C			

CR

iii.	Battery Backup (maintenance-free) Minimum 60 minutes on full charge.	C				
iv.	Should be provided with AC adaptor (3-pin square hospital grade plug top) and cigarette lighter socket cable for use in ambulance.	C				
3	All items supplied should be in-built in ambulance	C				
4	<b>Warranty, Services Agreements and Documentation</b>					
i.	All items described in this schedule should be covered by a warranty for a period of not less than 24 calendar months from the date of successful commissioning of the equipment including for any defect in manufacturing process, workmanship and all accessories	C				
ii.	A Labour-only service contract for a 03 year period and comprehensive maintenance contract (CMC) amount for thr whole equipment including installation material for another 5 years should be quoted as per schedule given below after 02 year warranty period a) 1 <sup>st</sup> & 2 <sup>nd</sup> year : warranty period b) 3 <sup>rd</sup> to 5 <sup>th</sup> year : Labour only contract c) 6 <sup>th</sup> to 10 <sup>th</sup> year : Comprehensive maintenance contract	C				
iii.	The technology manuals having full details of installation, operations, service and maintenance in the English language must be made available in duplicate at the time of installation and commissioning	C				
iv.	Spares/replaceable items/consumables required for the system must be guarantees for supply over a period of 10 years from the date of installation. The price list for such items over a period of 5 years from the date of supply should be made available	C				
v.	A sample unit must be available for demonstration and performance evaluation at the request of TEC	C				

vi. Prospective suppliers must provide documentary evidence of their previous experience with similar units installed in Sri Lanka as well as evidence of the reliability of after-sales services provided.	C				
vii. Training should be provided to end users in respect of operation, handling and maintenance of equipment supplied	C				

Make & Model: .....

Country of Manufacture: .....

Relevant catalogues & technical information attached.

Information given under "Bidder's Offer" is true and correct.

Bidder's Name, signature & Date

Company seal

- <sup>1</sup> Bidder shall fill in Col. 4, 5& 6
- <sup>2</sup> Purchaser will mark C for critical requirements
- <sup>3</sup> If the bidder has marked No for Conformity, bidder shall provide information in this column of his offer against the clause concerned in col. 2 (Purchaser's requirements). If the bidder has marked Yes, Bidder has the option of providing additional information to establish that it conforms to the specification given
- <sup>4</sup> Bidder shall include proof to support the conformity of the specification e.g. document and the page number in the original user/service manual mentioning the specification.

Technical Specifications

IN-BUILT MEDICAL ACCESSORIES FOR 02 NO'S AMBULANCE

Description: 02 No's Oxygen flow meter with regulator

1	2	3	4	5	6
	<b>KDU Requirements</b>				<b>Bidder's Offer</b>
		<b>Priority<sup>2</sup></b>	<b>Conformity</b>	<b>Proof<sup>4</sup></b>	<b>Remarks<sup>3</sup></b>
			Yes	No	
1	<b>Oxygen flow meter with regulator</b>				
	i. Adjustable flow rate -0-15 Lit/min	C			
	ii. Outlet pressure – 400kpa	C			
	iii. Regulator – Fixed output pressure diaphragm style	C			
	iv. Inlet connector – Pin-indexed yoke bull nosed	C			
	v. Outlet connector – Oxygen ring index	C			
	vi. Safety-Pressure Valve – Ball and seat type	C			
	vii. Bull nose type flow meter connector probe	C			
2	All items supplied should be in-built in ambulance	C			
3	<b>Warranty, Services Agreements and Documentation</b>				
	i. All items described in this schedule should be covered by a warranty for a period of not less than 24 calendar months from the date of successful commissioning of the equipment including for any defect in manufacturing process, workmanship and all accessories	C			

67

ii.	A Labour-only service contract for a 03 year period and comprehensive maintenance contract (CMC) amount for thr whole equipment including installation material for another 5 years should be quoted as per schedule given below after 02 year warranty period	C				
	a) 1 <sup>st</sup> & 2 <sup>nd</sup> year : warranty period					
	b) 3 <sup>rd</sup> to 5 <sup>th</sup> year : Labour only contract	C				
	c) 6 <sup>th</sup> to 10 <sup>th</sup> year : Comprehensive maintenance contract	C				
	iii. The technology manuals having full details of installation, operations, service and maintenance in the English language must be made available in duplicate at the time of installation and commissioning	C				
	iv. Spares/replaceable items/consumables required for the system must be guarantees for supply over a period of 10 years from the date of installation. The price list for such items over a period of 5 years from the date of supply should be made available	C				
	v. A sample unit must be available for demonstration and performance evaluation at the request of TEC	C				
	vi. Prospective suppliers must provide documentary evidence of their previous experience with similar units installed in Sri Lanka as well as evidence of the reliability of after-sales services provided.	C				
	vii. Training should be provided to end users in respect of operation, handling and maintenance of equipment supplied	C				

Make & Model: .....

Country of Manufacture: .....

Relevant catalogues & technical information attached.  
Information given under "Bidder's Offer" is true and correct.

Bidder's Name, signature & Date

Company seal

- <sup>1</sup> Bidder shall fill in Col. 4, 5& 6
- <sup>2</sup> Purchaser will mark C for critical requirements
- <sup>3</sup> If the bidder has marked No for Conformity, bidder shall provide information in this column of his offer against the clause concerned in col. 2 (Purchaser's requirements). If the bidder has marked Yes, Bidder has the option of providing additional information to establish that it conforms to the specification given
- <sup>4</sup> Bidder shall include proof to support the conformity of the specification e.g. document and the page number in the original user/service manual mentioning the specification.



**SPECIMEN FORM OF BID SECURITY**

By this Bond we ..... (hereinafter called "the Bidder") and We (name of bank or insurance company) whose registered office is at ..... (hereinafter called "the Surety") are held and firmly bound onto ..... (hereinafter called the Authority") in the sum of ..... for the payment of which sum the Bidder and the Surety bind themselves their successors and assigns jointly and severally by those presents.

Whereas the authority has invited the Tender and other persons to compete tenders in similar terms for the supply of ..... and to submit the same for the consideration of the Authority, and the Bidder proposes to submit to the Authority a Bid (hereafter called "the Bid") in accordance with such invitation, the Bond shall provide security to the Authority that the Bidder will honour certain obligations to be undertaken by him in the Tender in accordance with the following conditions.

**Now the Conditions of this Bond are:**

- (a) That it shall remain in full force and effect until the earliest of
  - (i) (Date), being ( ) days from (submission date), the date stipulated by the Authority for the submission of tenders, or any prolongation of such date above notified to the Authority by the Bidder and the Surety in writing.
  - (ii) In the event of acceptance of the Tender by the Authority, the date upon which the Bidder provides a performance security to the Authority in accordance with the terms of the contract thereby made between them, or
- (b) Subject to this Bond being in full force and effect, the Surety shall pay the full amount specified in this Bond upon receipt of first written demand form the Authority stating that.
  - (i) The Bidder has withdrawn his Tender during the validity of this Bond, or
  - (ii) The Bidder has failed to provide a performance security to the Authority in accordance with the terms of the tender within 14 days from receipt of intimation of award of the Tender.

No alteration in the terms of the Tender, nor any forbearance of forgiveness in or in respect of neither any matter or thing concerning the Tender on the part of the Authority, nor any objection from the bidder shall in any way release the Surety from any liability under this Bond.

The benefit of this Bond shall not be assignable by the Authority and upon its ceasing to be in full force and effect the Authority shall return the same to the Bidder.

This Bond shall be governed by the laws of Sri Lanka

I executed as a deed on this ( ) day of ( ) 20 ( )

For and on behalf of the Bidder

For and on behalf of the Surety

-----  
-----

-----  
-----

Signed by -----

Signed by -----

In the capacity of -----

In the capacity of -----

and by -----

and by -----

In the capacity of -----

In the capacity of -----

Seal (where applicable)

Seal (where applicable)

**DELIVERY SCHEDULE**

(IT IS MANDATORY TO FILL THE FOLLOWING SCHEDULE BY THE BIDDER) TENDER NO :

ITEM : ..... QTY :

DURATION	QTY
EX STOCK QTY (WITHIN 01 WEEK)	
01MONTH	
02 MONTHS	
03 MONTHS	
04 MONTHS	
<b>TOTAL</b>	

NAME OF THE BIDDER : .....

SIGNATURE OF BIDDER : .....

DATE : .....

COMPANY SEAL : .....